

CITO-SYSTEM GmbH

Haimendorfer Straße 37 + 46
90571 Schwaig bei Nürnberg
Deutschland

Test des Reinigers CITOCLYN Base der Firma CITO-SYSTEM GmbH

Sehr geehrte Damen und Herren,

Hiermit bestätigen wir, dass das Reinigungsmittel **CITOCLYN Base** der Firma CITO-SYSTEM GmbH in der uns übergebenen Form im Juni 2022 bei einem Langzeittest mit 54.000 Nutzungszyklen keine auffallenden Beschädigungen an den in unseren Anlagen verwendeten Stopp-Ventilen hervorgerufen hat. Dies entspricht unserem üblichen Testrahmen, der auch mit anderen Reinigern durchgeführt wurde. Somit kann das getestete Reinigungsmittel im Rahmen dieser Nutzungsdauer in unseren Anlagen **Final Plate Cleaner** und **Auto Clean System** verwendet werden.

Test of the cleaner CITOCLYN Base from CITO-SYSTEM GmbH

Dear Sirs,

We hereby confirm that the cleaning agent **CITOCLYN Base** from CITO-SYSTEM GmbH in the form handed over to us in June 2022 did not cause any noticeable damage to the stop valves used in our equipment in a long-term test with 54.000 usage cycles. This corresponds to our standard test framework, which was also carried out with other cleaning agents. This means that the tested cleaning agent can be used in our systems **Final Plate Cleaner** and **Auto-Clean System** within the above mentioned usage cycle.

Wartenberg, 15.09.2022


DUO-TECHNIK GmbH
Im Tiegel 4
36367 Wartenberg

Boris Kröll
Managing Director